



**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**
Esenyurt Firuzköy Bulvarı No:29 34325 Avcılar
İstanbul/ TÜRKİYE



AB-0583-T
20011163
03-20

TEST REPORT
DENEY RAPORU

Customer name: KURT KUMAŞ SAN.VE TİC.A.Ş
Address: 2. OSB 83228 Nolu CAD. NO:16 ŞEHİTKAMİL/GAZİANTEP
Buyer name: -
Contact Person: MURAT ERBAĞCI
Order No: -
Article No: -
Name and identity of test item: One sample of white nonwoven fabric. (Claimed to be; 25 GSM MELTBLOWN)
The date of receipt of test item: 19.03.2020
Re-submitted/re-confirmation date: -
Date of test: 19.03.2020-23.03.2020
Remarks: -
Sampling: The results given in this report belong to the received sample by vendor.
End-Use: -
Care Label: Not specified.
Number of pages of the report: 3

*The Turkish Accreditation Agency (TURKAK) is signatory to the multilateral agreements of the European co-operation for the Accreditation (EA) and of the International Laboratory Accreditation (ILAC) for the Mutual recognition of test reports.
EKOTEKS LABORATUVAR ve GÖZETİM HİZMETLERİ A.Ş. accredited by TÜRKAK under registration number [AB-0583-T] for ISO 17025:2017 as test laboratory.
The test and/or measurement results, the uncertainties (if applicable) with confidence probability and test methods are given on the following pages which are part of this report.*



Date
23.03.2020

Customer Representative
Ahmet ÇİRKİN

Head of Testing Laboratory
Sevim A. RAZAK
23.03.2020

This report shall not be reproduced other than in full except with the permission of the laboratory.
Testing reports without signature and seal are not valid.

EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.


AB-0583-T

20011163

03-20

REQUIRED TESTS	RESULT	COMMENTS
MICROBIOLOGICAL TEST		
Bacterial Filtration Efficiency (BFE)	-	
No requirements were given by vendor.		

REMARK: Original samples are kept for 3 months and all technical records are kept for 5 years unless otherwise specified. If requested, measurement uncertainty will be reported. But unless otherwise specified, measurement uncertainty is not considered while stating compliance with specification or limit values. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95 %. Tests marked (*) in this report are not included in the accreditation schedule.



This report shall not be reproduced other than in full except with the permission of the laboratory.
Testing reports without signature and seal are not val

AB-0583-T

20011163

03-20

TEST RESULTS

BACTERIAL FILTRATION EFFICIENCY (BFE)

Test Metod: EKOTEKS 70 (In-House Method-Bacterial Filtration Efficiency Testing /Ref: TS EN 14683:2019 Medical Face Masks, Requirements and Test Methods (*)

A specimen of the mask material is clamped between an impactor and an aerosol chamber. An aerosol of *Staphylococcus aureus* is introduced into the aerosol chamber and drawn through the mask material and the impactor under vacuum. The bacterial filtration efficiency of the mask is given by the number of colony forming units passing through the medical face mask material expressed as a percentage of the number of colony forming units present in the challenge aerosol.

Test Flow Rate	28,3 L/min
Test Flow Time	2 minute
Sample Sizes	20x20 cm ²
Microorganism	<i>Staphylococcus aureus</i> ATCC 6538
Bacterial concentration (cfu/ ml)	5x10 ⁵ cfu/ ml
incubation conditions	24 hour, 35°C ± 2°C
Positive control sample average of number of Bacteria (C)	2.6x10 ³ cfu/ ml

RESULTS		
Number of Test Sample	Test Sample (T) Number of Bacteria (cfu/ml)	Bacterial Filtration Efficiency (% B)
1	15	99.4 %
2	10	99.6 %
3	25	99.0 %
4	24	99.1 %
5	26	99.0 %

cfu: Colony-forming unit

$$B = (C - T) / C \times 100$$

%B: Bacterial Filtration Efficiency

C: is the mean of the total plate counts for the two positive control runs

T: is the total plate count for the test specimen



PF Eraser Free

EKOTEKS

EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.
Esenyurt Firuzköy Bulvarı No:29 34325 Avcılar
İstanbul/ TÜRKİYE

TEST REPORT
DENEY RAPORU

20009637

03-20

Customer name:

Address:

Buyer name:

Contact Person:

Order No:

Article No:

Name and Identity of test item:

The date of receipt of test item:

Re-submitted/re-confirmation
date:

Date of test:

Remarks:
Sampling:

End-Use:
Care Label:

Number of pages of the report: 4

MB - MELT BLOWN FABRIC

White non-woven fabric. (Claimed to be; 100 % PP , 30 gsm , Color Code:
White)

09.03.2020

-

09.03.2020-13.03.2020

-

The results given in this report belong to the received sample by vendor.

-

Not Specified

Gen.İT 36-2/03



Date
13.03.2020

Customer Representative
Hayriye ACARZELP

Head of Testing Laboratory
Sevim A. RAZAK
13.03.2020

This report shall not be reproduced other than in full except with the permission of the laboratory.
Testing reports without signature and seal are not valid.



EKOTEKS

**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**

20009637

03-20

REQUIRED TESTS	RESULT	COMMENTS
MICROBIOLOGICAL TEST		
Antibacterial Activity	P	
P: Pass F: Fail R: Refer to retailer technologist.		
Test results were evaluated according to EKOTEKS 70 (REF: 14683:2014) requirements		

REMARK: Original samples are kept for 3 months and all technical records are kept for 5 years unless otherwise specified. If requested, measurement uncertainty will be reported. But unless otherwise specified, measurement uncertainty is not considered while stating compliance with specification or limit values. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. Tests marked (*) in this report are not included in the accreditation schedule.

Gen.F136-03



This report shall not be reproduced other than in full except with the permission of the laboratory.
Testing reports without signature and seal are not valid.

20009537

03-20

TEST RESULT**BACTERIAL FILTRATION EFFICIENCY (BFE)**

Test Metod: EKOTEKS 70 (In-House Method-Bacterial Filtration Efficiency Testing /Ref: TS EN 14683:2019 Medical Face Masks, Requirements and Test Methods (*)

A specimen of the mask material is clamped between a impactor and an aerosol chamber. An aerosol of *Staphylococcus aureus* is introduced into the aerosol chamber and drawn through the mask material and the impactor under vacuum. The bacterial filtration efficiency of the mask is given by the number of colony forming units passing through the medical face mask material expressed as a percentage of the number of colony forming units present in the challenge aerosol.

Test Flow Rate	28,3 L/min
Test Flow Time	2 minute
Sample Sizes	5 piece (10x10 cm ²)
Microorganism	<i>Staphylococcus aureus</i> ATCC 6538
Bacterial concentration (cfu/ ml)	5x10 ⁷ cfu/ ml
Incubation conditions	24 hour, 35°C ± 2°C
Positive control average(C)	3.0x10 ⁷ cfu/ ml

Gen. P136-003

RESULTS		
Test Sample (T)	Number of Bacteria (cfu/ml)	Bacterial Filtration Efficiency (% B)
1	115	96.2 %
2	110	96.3 %
3	120	96.0 %
4	126	95.8 %
5	130	95.7 %

cfu: Colony-forming unit

$$B = (C - T) / C \times 100$$

%B: Bacterial Filtration Efficiency

C: is the mean of the total plate counts for the two positive control runs

T: is the total plate count for the test specimen



Results

No.	Test Item	Test Result
1	Bacterial Filtration Efficiency (BFE) Test	Specimen 1#: 99.9% Specimen 2#: 99.9% Specimen 3#: 99.8% Specimen 4#: 99.8% Specimen 5#: 99.7%
2	Differential Pressure Test	25.3 Pa/cm ²
3	Synthetic Blood Penetration Test	Specimen 1#~13#: None seen
4	Microbial Cleanliness Test	Specimen 1#: 22 CFU/g Specimen 2#: 14 CFU/g Specimen 3#: 10 CFU/g Specimen 4#: 8 CFU/g Specimen 5#: 10 CFU/g

Bacterial Filtration Efficiency (BFE) Test

1. Purpose

For evaluating the bacterial filtration efficiency (BFE) of mask.

2. Sample description was given by client

Sample description : Surgical masks
Specification : M
Lot Number : 200318
Sample Receiving Date : 2020-03-28

3. Test Method

EN 14683:2019+AC:2019(E) Annex B

4. Apparatus and materials

- 4.1 *Staphylococcus aureus* ATCC 6538.
- 4.2 Peptone water.
- 4.3 Tryptic Soy Broth(TSB).
- 4.4 Tryptic Soy Agar(TSA).
- 4.5 Bacterial filtration efficiency test apparatus.
- 4.6 Six-stage viable particle Anderson sampler.
- 4.7 Flow meters.

5. Test specimen

- 5.1 As requested by client, take a total of 5 test specimens.
- 5.2 Prior to testing, condition all test specimens for a minimum of 4 h at (21±5)°C and (85±5)% relative humidity.

TEST REPORT

Sample Description : Surgical masks
Sample Quantity : 50 pieces
Lot Number/Batch Code : 200318
Specification : M
Size : /
Type of Mask : Type IIR
Brand Name : /

Remark: The above information was provided by applicant.

Summary of Test Results

No.	Test Item	Test Standard	Judgement
1	Bacterial Filtration Efficiency (BFE) Test	EN 14683:2019+AC:2019(E) Annex B	Pass
2	Differential Pressure Test	EN 14683:2019+AC:2019(E) Annex C	Pass
3	Synthetic Blood Penetration Test	ISO 22609:2004	Pass
4	Microbial Cleanliness Test	EN 14683:2019+AC:2019(E) Annex D	Pass

Note: Pass = Meet customer requirements;
Fail = Fail customer requirements;
= No comment;
N.D. = Not detected.

Photo of Samples

